

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450,BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-300: BS-300, BS-320;

16.BS-1000M: BS-1000M, BS-1100M;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

18.BS-2800M:BS-2600M, BS-2800M;

330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

9.BS-380: BS-380, BS-390;

19.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

10.BS-400: BS-400, BS-420;

LOT : 061123006

2025-04-13

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа $\pm$ 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise $\pm$ 3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo $\pm$ 3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione $\pm$ 3 DS)
Türkçe	kisaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri $\pm$ 3SD)

Lp(a) II

English	Lp(a) II Calibrator
Русский	Калибратор Lp(a) II
Português	Calibrador Lp(a) II
Español	Calibrador Lp(a) II
Italiano	Calibratore Lp(a) II
Türkçe	Lp(a) II Kalibratörü

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value±3SD)			1 SD	Model	Assay Value	Range (Assay Value±3SD)			1 SD
Lp(a) II	nmol/L	BS-120 ¹	/	/	/	/	/	BS-120 ¹	/	/	/	/	/
		BS-200 ²	/	/	/	/	/	BS-200 ²	/	/	/	/	/
		BS-200E ³	/	/	/	/	/	BS-200E ³	/	/	/	/	/
		BS-240 ⁴	/	/	/	/	/	BS-240 ⁴	/	/	/	/	/
		BS-240E ⁵	/	/	/	/	/	BS-240E ⁵	/	/	/	/	/
		BS-300 ⁶	/	/	/	/	/	BS-300 ⁶	/	/	/	/	/
		BS-330E ⁷	/	/	/	/	/	BS-330E ⁷	/	/	/	/	/
		BS-360E ⁸	/	/	/	/	/	BS-360E ⁸	/	/	/	/	/
		BS-380 ⁹	/	/	/	/	/	BS-380 ⁹	/	/	/	/	/
		BS-400 ¹⁰	/	/	/	/	/	BS-400 ¹⁰	/	/	/	/	/
		BS-430 ¹¹	/	/	/	/	/	BS-430 ¹¹	/	/	/	/	/
		BS-480 ¹²	/	/	/	/	/	BS-480 ¹²	/	/	/	/	/
		BS-600 ¹³	/	/	/	/	/	BS-600 ¹³	/	/	/	/	/
		BS-600M ¹⁴	/	/	/	/	/	BS-600M ¹⁴	/	/	/	/	/

BS-800 ¹⁵	/	/	/	/	/	BS-800 ¹⁵	/	/	/	/	/
BS-1000M ¹⁶	43.1	28.7	—	57.5	4.8	BS-1000M ¹⁶	121	82	—	160	13
BS-2000 ¹⁷	/	/	/	/	/	BS-2000 ¹⁷	/	/	/	/	/
BS-2800M ¹⁸	/	/	/	/	/	BS-2800M ¹⁸	/	/	/	/	/

mindray

Lipoprotein (a) Control

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 061123006

 **2025-04-13**

