

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray

Português : Planilha de valores do Sistema de Medição BS da Mindray

Español : Hoja de valores del sistema de medición Mindray BS

Italiano : Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Português : A dados de cada grupo é a mesma.

Italiano : la dati di ogni gruppo è la stessa.

1.BS-200: BS-200, BS-220, BS-330, BS-350;

2.BS-200E: BS-200E, BS-220E,;

3.BS-240: BS-230, BS-240;

4.BS-240E: BS240E, BS240Pro;

5.BS-300: BS-300, BS-320;

6.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

7.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

8.BS-380: BS-380, BS-390;

9.BS-400: BS-400, BS-420;

Русский: Данные совпадают во всех группах.

Español : la datos de cada grupo es la misma.

Türkçe : her grubun veri aynıdır.

10.BS-430: BS-430, BS-450,BS-460;

11.BS-480: BS-480, BS-490;

12.BS-600: BS-600, BS-620;

13.BS-600M: BS-600M, BS-620M;

14.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

15.BS-1000M: BS-1000M, BS-1100M;

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

17.BS-2800M:BS-2600M, BS-2800M;

18.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

LOT

145124011

2025-09-27

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

CO2	
English	Carbon Dioxide
Русский	Двуокись углерода
Português	Dióxido de Carbono
Español	Dióxido de carbono
Italiano	Anidride carbonica
Türkçe	Karbon Dioksit

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)			1 SD	Model	Assay Value	Range(Assay Value±3SD)			1 SD
CO2	mmol/L	BS-200 ¹	13.2	10.5	—	15.9	0.9	BS-430 ¹⁰	13.5	10.8	—	16.2	0.9
		BS-200E ²	13.5	10.8	—	16.2	0.9	BS-480 ¹¹	13.5	10.8	—	16.2	0.9
		BS-240 ³	13.0	10.3	—	15.7	0.9	BS-600 ¹²	13.5	10.8	—	16.2	0.9
		BS-240E ⁴	13.5	10.8	—	16.2	0.9	BS-600M ¹³	14.0	11.3	—	16.7	0.9
		BS-300 ⁵	12.7	10.0	—	15.4	0.9	BS-800 ¹⁴	13.5	10.8	—	16.2	0.9
		BS-330E ⁶	13.5	10.8	—	16.2	0.9	BS-1000M ¹⁵	13.8	11.1	—	16.5	0.9
		BS-360E ⁷	13.5	10.8	—	16.2	0.9	BS-2000 ¹⁶	13.3	10.6	—	16.0	0.9
		BS-380 ⁸	13.5	10.8	—	16.2	0.9	BS-2800M ¹⁷	13.8	11.1	—	16.5	0.9
		BS-400 ⁹	13.5	10.8	—	16.2	0.9						
	mol/L	BS-200 ¹	0.0132	0.0105	—	0.0159	0.0009	BS-430 ¹⁰	0.0135	0.0108	—	0.0162	0.0009
		BS-200E ²	0.0135	0.0108	—	0.0162	0.0009	BS-480 ¹¹	0.0135	0.0108	—	0.0162	0.0009
		BS-240 ³	0.0130	0.0103	—	0.0157	0.0009	BS-600 ¹²	0.0135	0.0108	—	0.0162	0.0009
		BS-240E ⁴	0.0135	0.0108	—	0.0162	0.0009	BS-600M ¹³	0.0140	0.0113	—	0.0167	0.0009
		BS-300 ⁵	0.0127	0.0100	—	0.0154	0.0009	BS-800 ¹⁴	0.0135	0.0108	—	0.0162	0.0009

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)			1 SD	Model	Assay Value	Range(Assay Value±3SD)			1 SD
		BS-330E ⁶	0.0135	0.0108	—	0.0162	0.0009	BS-1000M ¹⁵	0.0138	0.0111	—	0.0165	0.0009
		BS-360E ⁷	0.0135	0.0108	—	0.0162	0.0009	BS-2000 ¹⁶	0.0133	0.0106	—	0.0160	0.0009
		BS-380 ⁸	0.0135	0.0108	—	0.0162	0.0009	BS-2800M ¹⁷	0.0138	0.0111	—	0.0165	0.0009
		BS-400 ⁹	0.0135	0.0108	—	0.0162	0.0009						

mindray

CO2 Control N

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 145124011

 **2025-09-27**

