

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
Português : Planilha de valores do Sistema de Medição BS da Mindray
Español : Hoja de valores del sistema de medición Mindray BS
Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Português : A dados de cada grupo é a mesma.

Italiano : la dati di ogni gruppo è la stessa.

1.BS-200: BS-200, BS-220, BS-330, BS-350;
2.BS-200E: BS-200E, BS-220E,;
3.BS-240: BS-230, BS-240;
4.BS-240E: BS240E, BS240Pro;
5.BS-300: BS-300, BS-320;
6.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")
7.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");
8.BS-380: BS-380, BS-390;
9.BS-400: BS-400, BS-420;

Русский: Данные совпадают во всех группах.

Español : la datos de cada grupo es la misma.

Türkçe : her grubun veri aynıdır.

10.BS-430: BS-430, BS-450,BS-460;
11.BS-480: BS-480, BS-490;
12.BS-600: BS-600, BS-620;
13.BS-600M: BS-600M, BS-620M;
14.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
15.BS-1000M: BS-1000M, BS-1100M;
16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
17.BS-2800M:BS-2600M, BS-2800M;
18.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

LOT

: 145124019

: 2026-02-18

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

CO2	
English	Carbon Dioxide
Русский	Двуокись углерода
Português	Dióxido de Carbono
Español	Dióxido de carbono
Italiano	Anidride carbonica
Türkçe	Karbon Dioksit

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)			1 SD	Model	Assay Value	Range(Assay Value±3SD)			1 SD
CO2	mmol/L	BS-200 ¹	12.4	10.0	—	14.8	0.8	BS-430 ¹⁰	12.8	10.1	—	15.5	0.9
		BS-200E ²	12.8	10.1	—	15.5	0.9	BS-480 ¹¹	12.8	10.1	—	15.5	0.9
		BS-240 ³	12.2	9.8	—	14.6	0.8	BS-600 ¹²	12.8	10.1	—	15.5	0.9
		BS-240E ⁴	12.8	10.1	—	15.5	0.9	BS-600M ¹³	13.0	10.3	—	15.7	0.9
		BS-300 ⁵	11.8	9.4	—	14.2	0.8	BS-800 ¹⁴	12.8	10.1	—	15.5	0.9
		BS-330E ⁶	12.8	10.1	—	15.5	0.9	BS-1000M ¹⁵	12.9	10.2	—	15.6	0.9
		BS-360E ⁷	12.8	10.1	—	15.5	0.9	BS-2000 ¹⁶	12.4	10.0	—	14.8	0.8
		BS-380 ⁸	12.8	10.1	—	15.5	0.9	BS-2800M ¹⁷	12.9	10.2	—	15.6	0.9
		BS-400 ⁹	12.8	10.1	—	15.5	0.9						
	mol/L	BS-200 ¹	0.0124	0.0100	—	0.0148	0.0008	BS-430 ¹⁰	0.0128	0.0101	—	0.0155	0.0009
		BS-200E ²	0.0128	0.0101	—	0.0155	0.0009	BS-480 ¹¹	0.0128	0.0101	—	0.0155	0.0009
		BS-240 ³	0.0122	0.0098	—	0.0146	0.0008	BS-600 ¹²	0.0128	0.0101	—	0.0155	0.0009
		BS-240E ⁴	0.0128	0.0101	—	0.0155	0.0009	BS-600M ¹³	0.0130	0.0103	—	0.0157	0.0009
		BS-300 ⁵	0.0118	0.0094	—	0.0142	0.0008	BS-800 ¹⁴	0.0128	0.0101	—	0.0155	0.0009

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)			1 SD	Model	Assay Value	Range(Assay Value±3SD)			1 SD
		BS-330E ⁶	0.0128	0.0101	—	0.0155	0.0009	BS-1000M ¹⁵	0.0129	0.0102	—	0.0156	0.0009
		BS-360E ⁷	0.0128	0.0101	—	0.0155	0.0009	BS-2000 ¹⁶	0.0124	0.0100	—	0.0148	0.0008
		BS-380 ⁸	0.0128	0.0101	—	0.0155	0.0009	BS-2800M ¹⁷	0.0129	0.0102	—	0.0156	0.0009
		BS-400 ⁹	0.0128	0.0101	—	0.0155	0.0009						

mindray

CO2 Control N

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 145124019

 **2026-02-18**

