

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
Português : Planilha de valores do Sistema de Medição BS da Mindray
Español : Hoja de valores del sistema de medición Mindray BS
Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.			Русский: Данные совпадают во всех группах.		
Português : A dados de cada grupo é a mesma.			Español : la datos de cada grupo es la misma.		
Italiano : la dati di ogni gruppo è la stessa.			Türkçe : her grubun veri aynıdır.		
1.BS-120: BS-120, BS-130, BS-180, BS-190;			11.BS-430: BS-430, BS-450,BS-460;		
2.BS-200: BS-200, BS-220, BS-330, BS-350;			12.BS-480: BS-480, BS-490;		
3.BS-200E: BS-200E, BS-220E,;			13.BS-600: BS-600, BS-620;		
4.BS-240: BS-230, BS-240;			14.BS-600M: BS-600M, BS-620M;		
5.BS-240E: BS240E, BS240Pro;			15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;		
6.BS-300: BS-300, BS-320;			16.BS-1000M: BS-1000M, BS-1100M;		
7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")			17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;		
8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");			18.BS-2800M:BS-2600M, BS-2800M;		
9.BS-380: BS-380, BS-390;			19.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.		
10.BS-400: BS-400, BS-420;					

LOT: 145324012 : 2025-12-14

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)
Fe					
English			Iron		
Русский			Железо		
Português			Ferro		
Español			Hierro		
Italiano			Ferro		
Türkçe			Demir		

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Assay Value	Range(Assay Value±3SD)	1 SD
Fe	μmol/L	BS-120 ¹	18.9	15.0 — 22.8	1.3	BS-400 ¹⁰	19.1	15.2 — 23.0	1.3
		BS-200 ²	18.9	15.0 — 22.8	1.3	BS-430 ¹¹	18.9	15.0 — 22.8	1.3
		BS-200E ³	19.2	15.3 — 23.1	1.3	BS-480 ¹²	19.1	15.2 — 23.0	1.3
		BS-240 ⁴	18.9	15.0 — 22.8	1.3	BS-600 ¹³	19.1	15.2 — 23.0	1.3
		BS-240E ⁵	18.8	14.9 — 22.7	1.3	BS-600M ¹⁴	18.8	14.9 — 22.7	1.3
		BS-300 ⁶	19.1	15.2 — 23.0	1.3	BS-800 ¹⁵	18.7	14.8 — 22.6	1.3
		BS-330E ⁷	19.2	15.3 — 23.1	1.3	BS-1000M ¹⁶	19.1	15.2 — 23.0	1.3
		BS-360E ⁸	19.2	15.3 — 23.1	1.3	BS-2000 ¹⁷	18.7	14.8 — 22.6	1.3
		BS-380 ⁹	19.1	15.2 — 23.0	1.3	BS-2800M ¹⁸	19.1	15.2 — 23.0	1.3
		BS-120 ¹	1.06	0.84 — 1.27	0.07	BS-400 ¹⁰	1.07	0.85 — 1.28	0.07
		BS-200 ²	1.06	0.84 — 1.27	0.07	BS-430 ¹¹	1.06	0.84 — 1.27	0.07
		BS-200E ³	1.07	0.85 — 1.29	0.07	BS-480 ¹²	1.07	0.85 — 1.28	0.07
		BS-240 ⁴	1.06	0.84 — 1.27	0.07	BS-600 ¹³	1.07	0.85 — 1.28	0.07

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)			1 SD	Model	Assay Value	Range(Assay Value±3SD)			1 SD
	mg/L	BS-240E ⁵	1.05	0.83	—	1.27	0.07	BS-600M ¹⁴	1.05	0.83	—	1.27	0.07
		BS-300 ⁶	1.07	0.85	—	1.28	0.07	BS-800 ¹⁵	1.04	0.83	—	1.26	0.07
		BS-330E ⁷	1.07	0.85	—	1.29	0.07	BS-1000M ¹⁶	1.07	0.85	—	1.28	0.07
		BS-360E ⁸	1.07	0.85	—	1.29	0.07	BS-2000 ¹⁷	1.04	0.83	—	1.26	0.07
		BS-380 ⁹	1.07	0.85	—	1.28	0.07	BS-2800M ¹⁸	1.07	0.85	—	1.28	0.07