

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
 Português : Planilha de valores do Sistema de Medição BS da Mindray
 Español : Hoja de valores del sistema de medición Mindray BS
 Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
 Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Português : A dados de cada grupo é a mesma.

Italiano : la dati di ogni gruppo è la stessa.

- 1.BS-120: BS-120, BS-130, BS-180, BS-190;
- 2.BS-200: BS-200, BS-220, BS-330, BS-350;
- 3.BS-200E: BS-200E, BS-220E,;
- 4.BS-240: BS-230, BS-240;
- 5.BS-240E: BS240E, BS240Pro;
- 6.BS-300: BS-300, BS-320;
- 7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")
- 8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");
- 9.BS-380: BS-380, BS-390;
- 10.BS-400: BS-400, BS-420;

Русский: Данные совпадают во всех группах.

Español : la datos de cada grupo es la misma.

Türkçe : her grubun veri aynıdır.

- 11.BS-430: BS-430, BS-450,BS-460;
- 12.BS-480: BS-480, BS-490;
- 13.BS-600: BS-600, BS-620;
- 14.BS-600M: BS-600M, BS-620M;
- 15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
- 16.BS-1000M: BS-1000M, BS-1100M;
- 17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
- 18.BS-2800M:BS-2600M, BS-2800M;
- 19.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

LOT : 145325008

有效期 : 2026-12-06

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

English		Fe
Iron		
Русский		Железо
Português		Ferro
Español		Hierro
Italiano		Ferro
Türkçe		Demir

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Assay Value	Range(Assay Value±3SD)	1 SD
Fe	µmol/L	BS-120¹	18.7	14.8 — 22.6	1.3	BS-400¹⁰	19.5	15.6 — 23.4	1.3
		BS-200²	18.7	14.8 — 22.6	1.3	BS-430¹¹	19.4	15.5 — 23.3	1.3
		BS-200E³	19.5	15.6 — 23.4	1.3	BS-480¹²	19.5	15.6 — 23.4	1.3
		BS-240⁴	18.7	14.8 — 22.6	1.3	BS-600¹³	19.5	15.6 — 23.4	1.3
		BS-240E⁵	19.3	15.4 — 23.2	1.3	BS-600M¹⁴	19.5	15.6 — 23.4	1.3
		BS-300⁶	19.5	15.6 — 23.4	1.3	BS-800¹⁵	19.2	15.3 — 23.1	1.3
		BS-330E⁷	19.5	15.6 — 23.4	1.3	BS-1000M¹⁶	19.5	15.6 — 23.4	1.3
		BS-360E⁸	19.5	15.6 — 23.4	1.3	BS-2000¹⁷	18.9	15.0 — 22.8	1.3
	BS-380⁹	19.5	15.6 — 23.4	1.3	BS-2800M¹⁸	19.5	15.6 — 23.4	1.3	
		BS-120¹	1.04	0.83 — 1.26	0.07	BS-400¹⁰	1.09	0.87 — 1.31	0.07
		BS-200²	1.04	0.83 — 1.26	0.07	BS-430¹¹	1.08	0.87 — 1.30	0.07
		BS-200E³	1.09	0.87 — 1.31	0.07	BS-480¹²	1.09	0.87 — 1.31	0.07
		BS-240⁴	1.04	0.83 — 1.26	0.07	BS-600¹³	1.09	0.87 — 1.31	0.07

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value \pm 3SD)			1 SD	Model	Assay Value	Range(Assay Value \pm 3SD)			1 SD
		BS-240E⁵	1.08	0.86	—	1.30	0.07	BS-600M¹⁴	1.09	0.87	—	1.31	0.07
		BS-300⁶	1.09	0.87	—	1.31	0.07	BS-800¹⁵	1.07	0.85	—	1.29	0.07
	mg/L	BS-330E⁷	1.09	0.87	—	1.31	0.07	BS-1000M¹⁶	1.09	0.87	—	1.31	0.07
		BS-360E⁸	1.09	0.87	—	1.31	0.07	BS-2000¹⁷	1.06	0.84	—	1.27	0.07
		BS-380⁹	1.09	0.87	—	1.31	0.07	BS-2800M¹⁸	1.09	0.87	—	1.31	0.07