

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray  
Português : Planilha de valores do Sistema de Medição BS da Mindray  
Español : Hoja de valores del sistema de medición Mindray BS  
Italiano : Scheda dei valori del sistema di misurazione BS di Mindray  
Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



**The data of each group is same.**

**Português : A dados de cada grupo é a mesma.**

**Italiano : la dati di ogni gruppo è la stessa.**

1.BS-120: BS-120, BS-130, BS-180, BS-190;  
2.BS-200: BS-200, BS-220, BS-330, BS-350;  
3.BS-200E: BS-200E, BS-220E,;  
4.BS-240: BS-230, BS-240;  
5.BS-240E: BS240E, BS240Pro;  
6.BS-300: BS-300, BS-320;  
7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")  
8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");  
9.BS-380: BS-380, BS-390;  
10.BS-400: BS-400, BS-420;

**Русский: Данные совпадают во всех группах.**

**Español : la datos de cada grupo es la misma.**

**Türkçe : her grubun veri aynıdır.**

11.BS-430: BS-430, BS-450,BS-460;  
12.BS-480: BS-480, BS-490;  
13.BS-600: BS-600, BS-620;  
14.BS-600M: BS-600M, BS-620M;  
15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;  
16.BS-1000M: BS-1000M, BS-1100M;  
17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;  
18.BS-2800M:BS-2600M, BS-2800M;  
19.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

LOT: 145325014

: 2027-03-28

| English   | Abbreviated name         | Model   | Unit    | Assay Value        | Range (Assay Value±3SD)                    |
|-----------|--------------------------|---------|---------|--------------------|--------------------------------------------|
| Русский   | сокращенное наименование | модель  | Прибор  | Результат анализа  | Диапазон(результат анализа ± 3CO)          |
| Português | Nome abreviado           | Modelo  | Unidade | Valores da análise | Faixa(Valores da análise ±3SD)             |
| Español   | nombre abreviado         | modelo  | Unidad  | Valor de ensayo    | Rango(Valor de ensayo ±3SD)                |
| Italiano  | abbreviazione            | modelli | Unità   | Valori di dosaggio | Intervallo (valore di concentrazione±3 DS) |
| Türkçe    | kısaltılmış ad           | model   | Ünite   | Tayin Değeri       | Aralık (Tayin Değeri±3SD)                  |

| Fe        |        |
|-----------|--------|
| English   | Iron   |
| Русский   | Железо |
| Português | Ferro  |
| Español   | Hierro |
| Italiano  | Ferro  |
| Türkçe    | Demir  |

| Abbreviated name | Unit   | Model                | Assay Value | Range(Assay Value±3SD) |   |      | 1 SD | Model                  | Assay Value | Range(Assay Value±3SD) |   |      | 1 SD |
|------------------|--------|----------------------|-------------|------------------------|---|------|------|------------------------|-------------|------------------------|---|------|------|
| Fe               | μmol/L | BS-120 <sup>1</sup>  | 19.8        | 15.9                   | — | 23.7 | 1.3  | BS-400 <sup>10</sup>   | 19.6        | 15.7                   | — | 23.5 | 1.3  |
|                  |        | BS-200 <sup>2</sup>  | 19.8        | 15.9                   | — | 23.7 | 1.3  | BS-430 <sup>11</sup>   | 19.2        | 15.3                   | — | 23.1 | 1.3  |
|                  |        | BS-200E <sup>3</sup> | 19.8        | 15.9                   | — | 23.7 | 1.3  | BS-480 <sup>12</sup>   | 19.6        | 15.7                   | — | 23.5 | 1.3  |
|                  |        | BS-240 <sup>4</sup>  | 19.8        | 15.9                   | — | 23.7 | 1.3  | BS-600 <sup>13</sup>   | 19.6        | 15.7                   | — | 23.5 | 1.3  |
|                  |        | BS-240E <sup>5</sup> | 19.4        | 15.5                   | — | 23.3 | 1.3  | BS-600M <sup>14</sup>  | 19.6        | 15.7                   | — | 23.5 | 1.3  |
|                  |        | BS-300 <sup>6</sup>  | 19.6        | 15.7                   | — | 23.5 | 1.3  | BS-800 <sup>15</sup>   | 19.6        | 15.7                   | — | 23.5 | 1.3  |
|                  |        | BS-330E <sup>7</sup> | 19.8        | 15.9                   | — | 23.7 | 1.3  | BS-1000M <sup>16</sup> | 19.6        | 15.7                   | — | 23.5 | 1.3  |
|                  |        | BS-360E <sup>8</sup> | 19.8        | 15.9                   | — | 23.7 | 1.3  | BS-2000 <sup>17</sup>  | 19.2        | 15.3                   | — | 23.1 | 1.3  |
|                  |        | BS-380 <sup>9</sup>  | 19.6        | 15.7                   | — | 23.5 | 1.3  | BS-2800M <sup>18</sup> | 19.6        | 15.7                   | — | 23.5 | 1.3  |
|                  |        | BS-120 <sup>1</sup>  | 1.11        | 0.89                   | — | 1.32 | 0.07 | BS-400 <sup>10</sup>   | 1.09        | 0.88                   | — | 1.31 | 0.07 |
|                  |        | BS-200 <sup>2</sup>  | 1.11        | 0.89                   | — | 1.32 | 0.07 | BS-430 <sup>11</sup>   | 1.07        | 0.85                   | — | 1.29 | 0.07 |
|                  |        | BS-200E <sup>3</sup> | 1.11        | 0.89                   | — | 1.32 | 0.07 | BS-480 <sup>12</sup>   | 1.09        | 0.88                   | — | 1.31 | 0.07 |
|                  |        | BS-240 <sup>4</sup>  | 1.11        | 0.89                   | — | 1.32 | 0.07 | BS-600 <sup>13</sup>   | 1.09        | 0.88                   | — | 1.31 | 0.07 |

| Abbreviated name | Unit | Model                | Assay Value | Range(Assay Value±3SD) |   |      | 1 SD | Model                  | Assay Value | Range(Assay Value±3SD) |   |      | 1 SD |
|------------------|------|----------------------|-------------|------------------------|---|------|------|------------------------|-------------|------------------------|---|------|------|
|                  | mg/L | BS-240E <sup>5</sup> | 1.08        | 0.87                   | — | 1.30 | 0.07 | BS-600M <sup>14</sup>  | 1.09        | 0.88                   | — | 1.31 | 0.07 |
|                  |      | BS-300 <sup>6</sup>  | 1.09        | 0.88                   | — | 1.31 | 0.07 | BS-800 <sup>15</sup>   | 1.09        | 0.88                   | — | 1.31 | 0.07 |
|                  |      | BS-330E <sup>7</sup> | 1.11        | 0.89                   | — | 1.32 | 0.07 | BS-1000M <sup>16</sup> | 1.09        | 0.88                   | — | 1.31 | 0.07 |
|                  |      | BS-360E <sup>8</sup> | 1.11        | 0.89                   | — | 1.32 | 0.07 | BS-2000 <sup>17</sup>  | 1.07        | 0.85                   | — | 1.29 | 0.07 |
|                  |      | BS-380 <sup>9</sup>  | 1.09        | 0.88                   | — | 1.31 | 0.07 | BS-2800M <sup>18</sup> | 1.09        | 0.88                   | — | 1.31 | 0.07 |