



## Reference Value Table

Product Name	Batch code And Use-by date	Abbreviation/Analyte	Applicable Instruments	Unit	Reference Value	Scope
D-Dimer control (low-value)	LOT <b>6224020216</b>  <b>2026-05-17</b>	D-D	CX Series Auto Coagulation Analyzers of Shenzhen Mindray Bio-medical Electronics Co., Ltd	µg/mL	0.35	0.18 ~ 0.53
				ng/mL	350	175 ~ 525

Product Name	Batch code And Use-by date	Abbreviation/Analyte	Applicable Instruments	Unit	Reference Value	Scope
D-Dimer control (low-value)	LOT <b>6224020216</b>  <b>2026-05-17</b>	D-D	C3510 and ExC810 Auto Coagulation Analyzers of Beijing Mindray Medical Instrument Co., Ltd	µg/mL	0.48	0.24 ~ 0.72
				ng/mL	480	240 ~ 720