


# Fibrin(ogen) Degradation Products (FDP) Control

## Reference Value Table

Product Name	Batch code And Use-by date	Abbreviation/ Analyte	Applicable Instruments	Unit	Reference Value	Scope
FDP control (low-value)	LOT 6624110116  2026-02-26	FDP	CX Series Auto Coagulation Analyzers of Shenzhen Mindray Bio-medical Electronics Co., Ltd	µg/mL	6.50	3.25 ~ 9.75
				mg/L	6.50	3.25 ~ 9.75