

Fibrin(ogen) Degradation Products (FDP) Control



Reference Value Table

Product Name	Batch code And Use-by date	Abbreviation/ Analyte	Applicable Instruments	Unit	Reference Value	Scope		
FDP control (high-value)	LOT 6824110116  2026-02-26	FDP	CX Series Auto Coagulation Analyzers of Shenzhen Mindray Bio-medical Electronics Co., Ltd	μg/mL	35.77	28.62	~	42.92
	mg/L			35.77	28.62	~	42.92	