

Reference Value Table

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation /Analyte	Unit	Reference Value	Scope
Coagulation control (level 1)	LOT	CX series Auto Coagulation Analyzers of Shenzhen Mindray Bio-medical Electronics Co., Ltd	PT	s	11.10	8.88 ~ 13.32
	8624050118		APTT	s	29.02	23.22 ~ 34.82
			TT	s	15.94	12.75 ~ 19.13
	2025-08-19		FIB	g/L	2.92	2.34 ~ 3.50
				mg/dL	292	234 ~ 350

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation /Analyte	Unit	Reference Value	Scope
Coagulation control (level 1)	LOT	C3510 and ExC810 Auto Coagulation Analyzers of Beijing Mindray Medical Instrument Co., Ltd	PT	s	13.06	10.45 ~ 15.67
	8624050118		APTT	s	29.35	23.48 ~ 35.22
			TT	s	14.65	11.72 ~ 17.58
	2025-08-19		FIB	g/L	2.79	2.23 ~ 3.35
				mg/dL	279	223 ~ 335