

Reference Value Table

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation /Analyte	Unit	Reference Value	Scope
Coagulation control (level 2)	LOT	CX series Auto Coagulation Analyzers of Shenzhen Mindray Bio-medical Electronics Co., Ltd	PT	s	38.62	30.90 ~ 46.34
	8725020218		APTT	s	59.70	47.76 ~ 71.64
			TT	s	/	/
	2026-05-18		FIB	g/L	/	/
				mg/dL	/	/

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation /Analyte	Unit	Reference Value	Scope
Coagulation control (level 2)	LOT	C3510 and ExC810 Auto Coagulation Analyzers of Beijing Mindray Medical Instrument Co., Ltd	PT	s	32.78	26.22 ~ 39.34
	8725020218		APTT	s	65.09	52.07 ~ 78.11
			TT	s	/	/
	2026-05-18		FIB	g/L	/	/
				mg/dL	/	/