


Low

LOT BC60624L-1

 2024-11-04

Parameter	Units	BC-6600, BC-6800		BC-6000, BC-6100, BC-6200, BC-6000Plus, BC-6100Plus		BC-6800Plus, BC-6700Plus, BC-6600Plus		BC-7500[N] CRP, BC-7500[NR] CRP, BC-7300[N] CRP, BC-7300[NR] CRP, BC-7500[N] CS, BC-7500[NR] CS		BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS		BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS	
		Values	Range	Values	Range	Values	Range	Values	Range	Values	Range	Values	Range
WBC	×10 <sup>9</sup> /L	2.78	± 0.80	2.74	± 0.80	2.96	± 0.80	2.70	± 0.80	2.82	± 0.80	2.82	± 0.80
Neu#	×10 <sup>9</sup> /L	1.84	± 0.50	1.77	± 0.50	1.94	± 0.50	1.75	± 0.50	1.93	± 0.50	1.93	± 0.50
Lym#	×10 <sup>9</sup> /L	0.62	± 0.40	0.64	± 0.40	0.67	± 0.40	0.63	± 0.40	0.61	± 0.40	0.61	± 0.40
Mon#	×10 <sup>9</sup> /L	0.18	± 0.12	0.18	± 0.12	0.19	± 0.12	0.17	± 0.12	0.17	± 0.12	0.17	± 0.12
Eos#	×10 <sup>9</sup> /L	0.11	± 0.30	0.11	± 0.30	0.12	± 0.30	0.11	± 0.30	0.07	± 0.30	0.07	± 0.30
Bas#	×10 <sup>9</sup> /L	0.03	± 0.15	0.04	± 0.15	0.04	± 0.15	0.04	± 0.15	0.04	± 0.15	0.04	± 0.15
IMG#*	×10 <sup>9</sup> /L	0.04	± 0.20	0.04	± 0.20	0.04	± 0.20	0.04	± 0.20	0.04	± 0.20	0.04	± 0.20
Neu%	%	66.3	± 12.0	64.8	± 12.0	65.6	± 12.0	64.8	± 12.0	68.2	± 12.0	68.2	± 12.0
Lym%	%	22.3	± 9.5	23.4	± 9.5	22.7	± 9.5	23.5	± 9.5	21.8	± 9.5	21.8	± 9.5
Mon%	%	6.3	± 3.0	6.5	± 3.0	6.3	± 3.0	6.4	± 3.0	6.2	± 3.0	6.2	± 3.0
Eos%	%	4.1	± 6.0	3.9	± 6.0	4.0	± 6.0	3.9	± 6.0	2.4	± 6.0	2.4	± 6.0
Bas%	%	1.0	± 5.0	1.4	± 5.0	1.4	± 5.0	1.4	± 5.0	1.4	± 5.0	1.4	± 5.0
IMG%*	%	1.3	± 5.0	1.3	± 5.0	1.3	± 5.0	1.3	± 5.0	1.3	± 5.0	1.3	± 5.0
RBC	×10 <sup>12</sup> /L	2.78	± 0.15	2.81	± 0.15	2.78	± 0.15	2.75	± 0.15	2.74	± 0.15	2.74	± 0.15
HGB	g/L	84	± 4	85	± 4	84	± 4	84	± 4	83	± 4	83	± 4
HCT	%	25.1	± 2.0	26.0	± 2.0	25.0	± 2.0	25.1	± 2.0	25.3	± 2.0	25.3	± 2.0
MCV	fL	90.3	± 5.0	92.6	± 5.0	90.1	± 5.0	91.2	± 5.0	92.4	± 5.0	92.4	± 5.0
MCH	pg	30.2	± 2.5	30.2	± 2.5	30.2	± 2.5	30.5	± 2.5	30.3	± 2.5	30.3	± 2.5
MCHC	g/L	335	± 30	327	± 30	335	± 30	335	± 30	328	± 30	328	± 30
RDW-CV	%	15.8	± 5.0	16.4	± 5.0	15.8	± 5.0	16.4	± 5.0	16.4	± 5.0	16.4	± 5.0
RDW-SD	fL	51.8	± 10.0	54.9	± 10.0	51.6	± 10.0	55.9	± 10.0	56.9	± 10.0	56.9	± 10.0
PLT	×10 <sup>9</sup> /L	82	± 20	85	± 20	78	± 20	82	± 20	82	± 20	82	± 20
PLT-H <sup>^^</sup>	×10 <sup>9</sup> /L	/	± /	/	± /	78	± 30	/	± /	83	± 30	83	± 30
PLT-O <sup>^</sup>	×10 <sup>9</sup> /L	/	± /	85	± 30	87	± 30	/	± /	/	± /	90	± 30
MPV	fL	8.3	± 3.0	9.2	± 3.0	8.2	± 3.0	8.8	± 3.0	8.7	± 3.0	8.7	± 3.0
PDW	/	13.7	± 5.0	14.0	± 5.0	13.6	± 5.0	13.7	± 5.0	13.6	± 5.0	13.6	± 5.0
PCT	%	0.068	± 0.030	0.078	± 0.030	0.064	± 0.030	0.072	± 0.030	0.071	± 0.030	0.071	± 0.030
P-LCR	%	6.8	± 10.0	10.9	± 10.0	6.0	± 10.0	8.2	± 10.0	6.6	± 10.0	6.6	± 10.0
P-LCC	×10 <sup>9</sup> /L	6	± 8	9	± 8	5	± 8	7	± 8	5	± 8	5	± 8
IPF**	%	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0	4.3	± 15.0	4.2	± 15.0	4.2	± 15.0
RET#****	×10 <sup>12</sup> /L	0.0156	± 0.0230	0.0230	± 0.0230	0.0239	± 0.0230	0.0215	± 0.0230	/	± /	0.0219	± 0.0230
RET%****	%	0.56	± 0.50	0.82	± 0.50	0.86	± 0.50	0.78	± 0.50	/	± /	0.80	± 0.50
IRF****	%	7.0	± 15.0	6.9	± 15.0	8.1	± 15.0	7.2	± 15.0	/	± /	7.6	± 15.0
LFRR****	%	93.0	± 15.0	93.1	± 15.0	91.9	± 15.0	92.8	± 15.0	/	± /	92.4	± 15.0
MFR****	%	4.6	± 13.0	4.5	± 13.0	5.7	± 13.0	4.9	± 13.0	/	± /	5.5	± 13.0
HFR****	%	2.4	± 2.0	2.4	± 2.0	2.4	± 2.0	2.3	± 2.0	/	± /	2.1	± 2.0
RHE***	pg	28.3	± 5.0	25.0	± 5.0	26.9	± 5.0	26.0	± 5.0	/	± /	26.2	± 5.0
NRBC#	×10 <sup>9</sup> /L	0.101	± 0.090	0.097	± 0.090	0.096	± 0.090	0.096	± 0.090	0.053	± 0.090	0.053	± 0.090
NRBC%	/100WBC	3.63	± 4.00	3.53	± 4.00	3.26	± 4.00	3.54	± 4.00	1.87	± 4.00	1.87	± 4.00
MCV <sup>r^^</sup>	fL	/	± /	106.4	± 19.0	112.8	± 19.0	/	± /	/	± /	/	± /
MCH <sup>r^^</sup>	pg	/	± /	29.0	± 6.3	30.8	± 6.3	/	± /	/	± /	/	± /
HDW <sup>^^^</sup>	g/L	/	± /	24.6	± 14.0	24.8	± 14.0	/	± /	/	± /	/	± /
DELTA-CH <sup>^^^</sup>	pg	/	± /	6.3	± 4.8	6.7	± 4.8	/	± /	/	± /	/	± /
HYPO-CH% <sup>^^^</sup>	%	/	± /	2.1	± 12.0	2.1	± 12.0	/	± /	/	± /	/	± /
HYPER-CH% <sup>^^^</sup>	%	/	± /	4.7	± 3.7	4.8	± 3.7	/	± /	/	± /	/	± /
MCHm <sup>^^^</sup>	pg	/	± /	22.7	± 5.0	24.2	± 5.0	/	± /	/	± /	/	± /
IPF <sup>^^^</sup>	×10 <sup>9</sup> /L	/	± /	3.5	± 8.0	3.3	± 8.0	/	± /	/	± /	/	± /
Micro <sup>^^^</sup>	%	/	± /	2.3	± 7.0	2.3	± 7.0	/	± /	/	± /	/	± /
Macro <sup>^^^</sup>	%	/	± /	5.8	± 8.0	5.1	± 8.0	/	± /	/	± /	/	± /
Neu-X <sup>^^^</sup>	/	/	± /	465.2	± 120.0	448.8	± 120.0	/	± /	/	± /	/	± /
Neu-Y <sup>^^^</sup>	/	/	± /	471.2	± 170.0	471.0	± 170.0	/	± /	/	± /	/	± /
Mon-X <sup>^^^</sup>	/	/	± /	305.3	± 120.0	273.7	± 120.0	/	± /	/	± /	/	± /
PLT-I <sup>^^^</sup>	×10 <sup>9</sup> /L	/	± /	85	± 30	78	± 30	/	± /	/	± /	/	± /

\*Applies to BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[N] CRP, BC-7500[NR] CRP, BC-7300[N] CRP, BC-7300[NR] CRP, BC-7500[N] CS, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and Software version 1.09 or higher of BC-6600/BC-6800.

\*\*Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and software version 1.09 or higher of BC-6600/BC-6800.

\*\*\*Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and Software version 1.10 or higher of BC-6800, BC-6600.

\*\*\*\*Applies to BC-6600, BC-6800, BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

\*Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

\*\*Applies to BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS, BC-6600Plus, BC-6700Plus, BC-6800Plus.

\*\*\*Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus.

Before using, refer to the instruction sheet for mixing directions.



Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China  
Tel: +86 755 26582888  
Fax: +86 755 26582880

EC   REF

Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80 20537 Hamburg, Germany  
Tel: 0049-40-2513175  
Fax: 0049-40-255726

IVD

P/N: 046-008311-00 (10.0)

Normal		[LOT] BC60624N-1		📅 2024-11-04									
Parameter	Units	BC-6600, BC-6800		BC-6000,BC-6100, BC-6200, BC-6000Plus, BC-6100Plus		BC-6800Plus, BC-6700Plus, BC-6600Plus		BC-7500[N] CRP, BC-7500[NR] CRP, BC-7300[N] CRP, BC-7300[NR] CRP, BC-7500[N] CS, BC-7500[NR] CS		BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS		BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS	
		Values	Range	Values	Range	Values	Range	Values	Range	Values	Range	Values	Range
WBC	×10 <sup>9</sup> /L	8.74	± 1.00	8.74	± 1.00	9.03	± 1.00	8.50	± 1.00	8.96	± 1.00	8.96	± 1.00
Neu#	×10 <sup>9</sup> /L	5.76	± 0.80	5.71	± 0.80	5.91	± 0.80	5.53	± 0.80	6.10	± 0.80	6.10	± 0.80
Lym#	×10 <sup>9</sup> /L	2.00	± 0.80	2.01	± 0.80	2.07	± 0.80	1.97	± 0.80	1.95	± 0.80	1.95	± 0.80
Mon#	×10 <sup>9</sup> /L	0.54	± 0.38	0.55	± 0.38	0.56	± 0.38	0.54	± 0.38	0.56	± 0.38	0.56	± 0.38
Eos#	×10 <sup>9</sup> /L	0.35	± 0.50	0.35	± 0.50	0.36	± 0.50	0.34	± 0.50	0.22	± 0.50	0.22	± 0.50
Bas#	×10 <sup>9</sup> /L	0.09	± 0.50	0.12	± 0.50	0.13	± 0.50	0.12	± 0.50	0.13	± 0.50	0.13	± 0.50
IMG#*	×10 <sup>9</sup> /L	0.10	± 0.40	0.11	± 0.40	0.12	± 0.40	0.10	± 0.40	0.12	± 0.40	0.12	± 0.40
Neu%	%	65.9	± 10.0	65.3	± 10.0	65.5	± 10.0	65.0	± 10.0	68.1	± 10.0	68.1	± 10.0
Lym%	%	22.9	± 10.0	23.0	± 10.0	22.9	± 10.0	23.2	± 10.0	21.8	± 10.0	21.8	± 10.0
Mon%	%	6.2	± 4.0	6.3	± 4.0	6.2	± 4.0	6.4	± 4.0	6.2	± 4.0	6.2	± 4.0
Eos%	%	4.0	± 6.0	4.0	± 6.0	4.0	± 6.0	4.0	± 6.0	2.5	± 6.0	2.5	± 6.0
Bas%	%	1.0	± 5.0	1.4	± 5.0	1.4	± 5.0	1.4	± 5.0	1.4	± 5.0	1.4	± 5.0
IMG%*	%	1.2	± 5.0	1.3	± 5.0	1.3	± 5.0	1.2	± 5.0	1.3	± 5.0	1.3	± 5.0
RBC	×10 <sup>12</sup> /L	4.47	± 0.24	4.54	± 0.24	4.52	± 0.24	4.53	± 0.24	4.49	± 0.24	4.49	± 0.24
HGB	g/L	137	± 5	138	± 5	137	± 5	138	± 5	137	± 5	137	± 5
HCT	%	41.1	± 2.5	42.3	± 2.5	41.7	± 2.5	42.3	± 2.5	42.4	± 2.5	42.4	± 2.5
MCV	fL	92.0	± 5.0	93.2	± 5.0	92.2	± 5.0	93.4	± 5.0	94.5	± 5.0	94.5	± 5.0
MCH	pg	30.6	± 2.5	30.4	± 2.5	30.3	± 2.5	30.5	± 2.5	30.5	± 2.5	30.5	± 2.5
MCHC	g/L	333	± 30	326	± 30	329	± 30	326	± 30	323	± 30	323	± 30
RDW-CV	%	16.0	± 5.0	16.4	± 5.0	15.9	± 5.0	15.9	± 5.0	15.9	± 5.0	15.9	± 5.0
RDW-SD	fL	53.3	± 10.0	55.4	± 10.0	52.9	± 10.0	55.2	± 10.0	55.9	± 10.0	55.9	± 10.0
PLT	×10 <sup>9</sup> /L	222	± 40	236	± 40	219	± 40	224	± 40	223	± 40	223	± 40
PLT-H^A	×10 <sup>9</sup> /L	/	± /	/	± /	218	± 60	/	± /	226	± 60	226	± 60
PLT-O^A	×10 <sup>9</sup> /L	/	± /	254	± 60	239	± 60	/	± /	/	± /	256	± 60
MPV	fL	8.3	± 3.0	9.1	± 3.0	8.3	± 3.0	8.9	± 3.0	8.7	± 3.0	8.7	± 3.0
PDW	/	13.7	± 5.0	14.1	± 5.0	13.7	± 5.0	13.8	± 5.0	13.6	± 5.0	13.6	± 5.0
PCT	%	0.184	± 0.070	0.215	± 0.070	0.182	± 0.070	0.199	± 0.070	0.194	± 0.070	0.194	± 0.070
P-LCR	%	7.2	± 10.0	11.8	± 10.0	7.4	± 10.0	9.9	± 10.0	8.2	± 10.0	8.2	± 10.0
P-LCC	×10 <sup>9</sup> /L	16	± 24	28	± 24	16	± 24	22	± 24	18	± 24	18	± 24
IPF**	%	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0
RET#****	×10 <sup>12</sup> /L	0.1064	± 0.0990	0.1221	± 0.0990	0.1234	± 0.0990	0.1191	± 0.0990	/	± /	0.1253	± 0.0990
RET%****	%	2.38	± 2.00	2.69	± 2.00	2.73	± 2.00	2.63	± 2.00	/	± /	2.79	± 2.00
IRF****	%	6.8	± 13.0	7.2	± 13.0	8.0	± 13.0	7.1	± 13.0	/	± /	7.4	± 13.0
LFR****	%	93.2	± 13.0	92.8	± 13.0	92.0	± 13.0	92.9	± 13.0	/	± /	92.6	± 13.0
MFR****	%	4.3	± 11.0	5.2	± 11.0	5.9	± 11.0	5.0	± 11.0	/	± /	5.5	± 11.0
HFR****	%	2.5	± 2.0	2.0	± 2.0	2.1	± 2.0	2.1	± 2.0	/	± /	1.9	± 2.0
RHE***	pg	28.3	± 5.0	24.2	± 5.0	25.7	± 5.0	25.1	± 5.0	/	± /	25.9	± 5.0
NRBC#	×10 <sup>9</sup> /L	0.300	± 0.270	0.311	± 0.270	0.303	± 0.270	0.301	± 0.270	0.170	± 0.270	0.170	± 0.270
NRBC%	/100WBC	3.43	± 4.00	3.56	± 4.00	3.36	± 4.00	3.54	± 4.00	1.90	± 4.00	1.90	± 4.00
MCV^AAA	fL	/	± /	107.1	± 13.0	113.8	± 13.0	/	± /	/	± /	/	± /
MCH^AAA	pg	/	± /	29.2	± 6.3	31.2	± 6.3	/	± /	/	± /	/	± /
HDW^AAA	g/L	/	± /	24.3	± 14.0	24.9	± 14.0	/	± /	/	± /	/	± /
DELTA-CH^AAA	pg	/	± /	6.3	± 4.8	6.8	± 4.8	/	± /	/	± /	/	± /
HYPO-CH^AAA	%	/	± /	2.2	± 14.0	2.2	± 14.0	/	± /	/	± /	/	± /
HYPER-CH^AAA	%	/	± /	4.7	± 4.0	4.6	± 4.0	/	± /	/	± /	/	± /
MCHm^AAA	pg	/	± /	22.9	± 5.0	24.4	± 5.0	/	± /	/	± /	/	± /
IPF#^AAA	×10 <sup>9</sup> /L	/	± /	9.8	± 25.0	9.2	± 25.0	/	± /	/	± /	/	± /
Micro^AAA	%	/	± /	2.3	± 3.0	2.3	± 3.0	/	± /	/	± /	/	± /
Macro^AAA	%	/	± /	5.9	± 10.0	5.2	± 10.0	/	± /	/	± /	/	± /
Neu-X^AAA	/	/	± /	475.2	± 120.0	447.2	± 120.0	/	± /	/	± /	/	± /
Neu-Y^AAA	/	/	± /	478.4	± 170.0	467.5	± 170.0	/	± /	/	± /	/	± /
Mon-X^AAA	/	/	± /	307.9	± 120.0	274.4	± 120.0	/	± /	/	± /	/	± /
PLT-J^AAA	×10 <sup>9</sup> /L	/	± /	236	± 60	219	± 60	/	± /	/	± /	/	± /

\*Applies to BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[N] CRP, BC-7500[NR] CRP, BC-7300[N] CRP, BC-7300[NR] CRP, BC-7500[N] CS, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and Software version 1.09 or higher of BC-6600/BC-6800.

\*\*Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and software version 1.09 or higher of BC-6600/BC-6800.

\*\*\*Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and Software version 1.10 or higher of BC-6800, BC-6600.

\*\*\*\*Applies to BC-6600, BC-6800, BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

\*Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

\*\*Applies to BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS, BC-6600Plus, BC-6700Plus, BC-6800Plus.

\*\*\*Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus.

Before using, refer to the instruction sheet for mixing directions.



Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China  
Tel: +86 755 26582888  
Fax: +86 755 26582680

EC REP

Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80 20537 Hamburg, Germany  
Tel: 0049-40-2513175  
Fax: 0049-40-255726

IVD

High

LOT BC60624H-1

2024-11-04

Parameter	Units	BC-6600, BC-6800		BC-6000,BC-6100, BC-6200, BC-6000Plus, BC-6100Plus		BC-6800Plus, BC-6700Plus, BC-6600Plus		BC-7500[N] CRP, BC-7500[NR] CRP, BC-7300[N] CRP, BC-7300[NR] CRP, BC-7500[N] CS, BC-7500[NR] CS		BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS		BC-700[R], BC-720[R], BC-760[R], BC-700[R] CS, BC-760[R] CS	
		Values	Range	Values	Range	Values	Range	Values	Range	Values	Range	Values	Range
WBC	×10 <sup>9</sup> /L	19.84	± 2.00	20.59	± 2.00	20.72	± 2.00	19.98	± 2.00	20.59	± 2.00	20.59	± 2.00
Neu#	×10 <sup>9</sup> /L	12.82	± 2.00	13.30	± 2.00	13.44	± 2.00	12.92	± 2.00	13.79	± 2.00	13.79	± 2.00
Lym#	×10 <sup>9</sup> /L	4.76	± 1.60	4.92	± 1.60	4.93	± 1.60	4.78	± 1.60	4.76	± 1.60	4.76	± 1.60
Mon#	×10 <sup>9</sup> /L	1.29	± 0.80	1.32	± 0.80	1.31	± 0.80	1.28	± 0.80	1.28	± 0.80	1.28	± 0.80
Eos#	×10 <sup>9</sup> /L	0.77	± 1.20	0.82	± 1.20	0.81	± 1.20	0.78	± 1.20	0.47	± 1.20	0.47	± 1.20
Bas#	×10 <sup>9</sup> /L	0.20	± 1.00	0.23	± 1.00	0.23	± 1.00	0.22	± 1.00	0.29	± 1.00	0.29	± 1.00
IMG#*	×10 <sup>9</sup> /L	0.26	± 1.20	0.25	± 1.20	0.25	± 1.20	0.26	± 1.20	0.27	± 1.20	0.27	± 1.20
Neu%	%	64.6	± 10.0	64.6	± 10.0	64.9	± 10.0	64.7	± 10.0	67.0	± 10.0	67.0	± 10.0
Lym%	%	24.0	± 8.0	23.9	± 8.0	23.8	± 8.0	23.9	± 8.0	23.1	± 8.0	23.1	± 8.0
Mon%	%	6.5	± 3.5	6.4	± 3.5	6.3	± 3.5	6.4	± 3.5	6.2	± 3.5	6.2	± 3.5
Eos%	%	3.9	± 6.0	4.0	± 6.0	3.9	± 6.0	3.9	± 6.0	2.3	± 6.0	2.3	± 6.0
Bas%	%	1.0	± 5.0	1.1	± 5.0	1.1	± 5.0	1.1	± 5.0	1.4	± 5.0	1.4	± 5.0
IMG%*	%	1.3	± 5.0	1.2	± 5.0	1.2	± 5.0	1.3	± 5.0	1.3	± 5.0	1.3	± 5.0
RBC	×10 <sup>12</sup> /L	5.19	± 0.30	5.28	± 0.30	5.20	± 0.30	5.29	± 0.30	5.25	± 0.30	5.25	± 0.30
HGB	g/L	160	± 6	162	± 6	161	± 6	163	± 6	162	± 6	162	± 6
HCT	%	48.1	± 3.0	49.4	± 3.0	48.5	± 3.0	49.9	± 3.0	50.2	± 3.0	50.2	± 3.0
MCV	fL	92.7	± 5.0	93.6	± 5.0	93.2	± 5.0	94.4	± 5.0	95.6	± 5.0	95.6	± 5.0
MCH	pg	30.8	± 2.5	30.7	± 2.5	31.0	± 2.5	30.8	± 2.5	30.9	± 2.5	30.9	± 2.5
MCHC	g/L	333	± 30	328	± 30	332	± 30	326	± 30	323	± 30	323	± 30
RDW-CV	%	16.2	± 6.0	16.3	± 6.0	16.1	± 6.0	15.8	± 6.0	15.8	± 6.0	15.8	± 6.0
RDW-SD	fL	54.3	± 12.0	55.4	± 12.0	53.8	± 12.0	55.1	± 12.0	56.0	± 12.0	56.0	± 12.0
PLT	×10 <sup>9</sup> /L	507	± 60	536	± 60	501	± 60	511	± 60	508	± 60	508	± 60
PLT-H <sup>^</sup>	×10 <sup>9</sup> /L	/	± /	/	± /	498	± 120	/	± /	515	± 120	515	± 120
PLT-O <sup>^</sup>	×10 <sup>9</sup> /L	/	± /	591	± 120	547	± 120	/	± /	/	± /	589	± 120
MPV	fL	8.4	± 3.0	9.0	± 3.0	8.4	± 3.0	8.9	± 3.0	8.8	± 3.0	8.8	± 3.0
PDW	/	13.7	± 5.0	14.0	± 5.0	13.7	± 5.0	13.7	± 5.0	13.6	± 5.0	13.6	± 5.0
PCT	%	0.426	± 0.110	0.482	± 0.110	0.421	± 0.110	0.455	± 0.110	0.447	± 0.110	0.447	± 0.110
P-LCR	%	7.6	± 10.0	11.3	± 10.0	7.6	± 10.0	10.1	± 10.0	8.6	± 10.0	8.6	± 10.0
P-LCC	×10 <sup>9</sup> /L	39	± 50	61	± 50	38	± 50	52	± 50	44	± 50	44	± 50
IPF**	%	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0	4.2	± 15.0
RET#****	×10 <sup>12</sup> /L	0.1925	± 0.1200	0.2123	± 0.1200	0.2106	± 0.1200	0.2090	± 0.1200	/	± /	0.2237	± 0.1200
RET%****	%	3.71	± 3.00	4.02	± 3.00	4.05	± 3.00	3.95	± 3.00	/	± /	4.26	± 3.00
IRF****	%	6.8	± 20.0	7.2	± 20.0	8.1	± 20.0	7.2	± 20.0	/	± /	7.3	± 20.0
LFR****	%	93.2	± 20.0	92.8	± 20.0	91.9	± 20.0	92.8	± 20.0	/	± /	92.7	± 20.0
MFR****	%	4.2	± 18.0	5.2	± 18.0	6.0	± 18.0	5.0	± 18.0	/	± /	5.3	± 18.0
HFR****	%	2.6	± 2.0	2.0	± 2.0	2.1	± 2.0	2.2	± 2.0	/	± /	2.0	± 2.0
RHE***	pg	28.5	± 5.0	24.5	± 5.0	26.0	± 5.0	25.4	± 5.0	/	± /	26.2	± 5.0
NRBC#	×10 <sup>9</sup> /L	0.710	± 0.600	0.745	± 0.600	0.719	± 0.600	0.709	± 0.600	0.379	± 0.600	0.379	± 0.600
NRBC%	/100WBC	3.58	± 4.00	3.62	± 4.00	3.47	± 4.00	3.55	± 4.00	1.84	± 4.00	1.84	± 4.00
MCV <sup>^</sup> ^^	fL	/	± /	107.6	± 13.0	114.4	± 13.0	/	± /	/	± /	/	± /
MCH <sup>^</sup> ^^	pg	/	± /	29.5	± 6.3	31.4	± 6.3	/	± /	/	± /	/	± /
HDW <sup>^^</sup>	g/L	/	± /	24.1	± 14.0	24.4	± 14.0	/	± /	/	± /	/	± /
DELTA-CH <sup>^^^</sup>	pg	/	± /	6.4	± 4.8	6.9	± 4.8	/	± /	/	± /	/	± /
HYPO-CH <sup>^^^</sup>	%	/	± /	2.1	± 14.0	2.2	± 14.0	/	± /	/	± /	/	± /
HYPER-CH <sup>^^^</sup>	%	/	± /	4.5	± 4.7	4.6	± 4.7	/	± /	/	± /	/	± /
MCHm <sup>^^^</sup>	pg	/	± /	23.0	± 5.0	24.6	± 5.0	/	± /	/	± /	/	± /
IPF# <sup>^^^</sup>	×10 <sup>9</sup> /L	/	± /	22.6	± 40.0	21.2	± 40.0	/	± /	/	± /	/	± /
Micro <sup>^^^</sup>	%	/	± /	2.4	± 5.0	2.2	± 5.0	/	± /	/	± /	/	± /
Macro <sup>^^^</sup>	%	/	± /	6.0	± 10.0	5.1	± 10.0	/	± /	/	± /	/	± /
Neu-X <sup>^^^</sup>	/	/	± /	462.0	± 120.0	445.0	± 120.0	/	± /	/	± /	/	± /
Neu-Y <sup>^^^</sup>	/	/	± /	462.1	± 170.0	461.8	± 170.0	/	± /	/	± /	/	± /
Mon-X <sup>^^^</sup>	/	/	± /	284.9	± 120.0	255.1	± 120.0	/	± /	/	± /	/	± /
PLT- <sup>^</sup> ^^	×10 <sup>9</sup> /L	/	± /	536	± 120	501	± 120	/	± /	/	± /	/	± /

\*Applies to BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[N] CRP, BC-7500[NR] CRP, BC-7300[N] CRP, BC-7300[NR] CRP, BC-7500[N] CS, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-700[R] CS and Software version 1.09 or higher of BC-6600/BC-6800.

\*\*Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and software version 1.09 or higher of BC-6600/BC-6800.

\*\*\*Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and Software version 1.10 or higher of BC-6800, BC-6600.

\*\*\*\*Applies to BC-6600, BC-6800, BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

\*Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

\*\*Applies to BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS, BC-6600Plus, BC-6700Plus, BC-6800Plus.

\*\*\*Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus.

Before using, refer to the instruction sheet for mixing directions.

  
Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China  
Tel: +86 755 26582888  
Fax: +86 755 26582680

EC REP

Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80 20537 Hamburg, Germany  
Tel: 0049-40-2513175  
Fax: 0049-40-255726

IVD

P/N: 046-008311-00 (10.0)