

LOT BC61024

2025-03-10

Instrument	Parameter	Unit	Low Level			Normal Level			High Level			
			LOT		BC61024L-1		LOT		BC61024N-1		LOT	
			Values	Range	Values	Range	Values	Range	Values	Range	Values	Range
BC-700[R] BC-720[R] BC-760[R] BC-780[R] BC-700[R] CS BC-760[R] CS	WBC	$\times 10^9/L$	2.99	±	0.80	9.02	±	1.00	20.39	±	2.00	
	Neu#	$\times 10^9/L$	2.01	±	0.50	6.07	±	0.80	13.70	±	2.00	
	Lym#	$\times 10^9/L$	0.68	±	0.40	2.05	±	0.80	4.67	±	1.60	
	Mon#	$\times 10^9/L$	0.19	±	0.12	0.56	±	0.38	1.26	±	0.80	
	Eos#	$\times 10^9/L$	0.07	±	0.30	0.21	±	0.50	0.47	±	1.20	
	Bas#	$\times 10^9/L$	0.04	±	0.15	0.13	±	0.50	0.29	±	1.00	
	IMG#*	$\times 10^9/L$	0.04	±	0.20	0.12	±	0.40	0.27	±	1.20	
	Neu%	%	67.0	±	12.0	67.4	±	10.0	67.2	±	10.0	
	Lym%	%	22.9	±	9.5	22.7	±	10.0	22.9	±	8.0	
	Mon%	%	6.4	±	3.0	6.2	±	4.0	6.2	±	3.5	
	Eos%	%	2.4	±	6.0	2.3	±	6.0	2.3	±	6.0	
	Bas%	%	1.3	±	5.0	1.4	±	5.0	1.4	±	5.0	
	IMG%*	%	1.3	±	5.0	1.3	±	5.0	1.3	±	5.0	
	RBC	$\times 10^{12}/L$	2.70	±	0.15	4.57	±	0.24	5.34	±	0.30	
	HGB	g/L	83	±	4	140	±	5	170	±	6	
	HCT	%	25.1	±	2.0	43.6	±	2.5	53.5	±	3.0	
	MCV	fL	93.0	±	5.0	95.5	±	5.0	100.2	±	5.0	
	MCH	pg	30.7	±	2.5	30.6	±	2.5	31.8	±	2.5	
	MCHC	g/L	331	±	30	321	±	30	318	±	30	
	RDW-CV	%	15.4	±	5.0	14.8	±	5.0	14.4	±	6.0	
	RDW-SD	fL	53.9	±	10.0	53.0	±	10.0	53.4	±	12.0	
	PLT	$\times 10^9/L$	102	±	20	241	±	40	512	±	60	
	PLT-H ^{MM}	$\times 10^9/L$	103	±	30	244	±	60	519	±	120	
	PLT-O ^A	$\times 10^9/L$	104	±	30	268	±	60	592	±	120	
	MPV	fL	9.8	±	3.0	9.9	±	3.0	9.8	±	3.0	
	PDW	/	13.7	±	5.0	13.7	±	5.0	13.8	±	5.0	
	PCT	%	0.100	±	0.030	0.239	±	0.070	0.502	±	0.110	
	P-LCR	%	14.3	±	10.0	15.7	±	10.0	16.4	±	10.0	
	P-LCC	$\times 10^9/L$	15	±	8	38	±	24	84	±	50	
	IPF**	%	4.2	±	15.0	4.2	±	15.0	4.2	±	15.0	
	Ret#****	$\times 10^{12}/L$	0.0308	±	0.0230	0.1133	±	0.0990	0.2179	±	0.1200	
	Ret%****	%	1.14	±	0.50	2.48	±	2.00	4.08	±	3.00	
	IRF****	%	12.1	±	15.0	15.6	±	13.0	15.3	±	20.0	
	LFR****	%	87.9	±	15.0	84.4	±	13.0	84.7	±	20.0	
	MFR****	%	10.1	±	13.0	13.6	±	11.0	13.3	±	18.0	
	HFR****	%	2.0	±	2.0	2.0	±	2.0	2.0	±	2.0	
	RHE***	pg	26.4	±	5.0	25.5	±	5.0	25.9	±	5.0	
	NRBC#	$\times 10^9/L$	0.056	±	0.090	0.169	±	0.270	0.375	±	0.600	
	NRBC%	/100WBC	1.86	±	4.00	1.87	±	4.00	1.84	±	4.00	

LOT BC61024

2025-03-10

Instrument	Parameter	Unit	Low Level			Normal Level			High Level			
			LOT		BC61024L-1		LOT		BC61024N-1		LOT	
			Values	Range	Values	Range	Values	Range	Values	Range	Values	Range
BC-7500[N] CRP BC-7500[NR] CRP BC-7300[N] CRP BC-7300[NR] CRP BC-7500[N] CS BC-7500[NR] CS	WBC	$\times 10^9/L$	2.96	±	0.80	8.82	±	1.00	19.99	±	2.00	
	Neu#	$\times 10^9/L$	1.91	±	0.50	5.77	±	0.80	13.05	±	2.00	
	Lym#	$\times 10^9/L$	0.70	±	0.40	2.03	±	0.80	4.66	±	1.60	
	Mon#	$\times 10^9/L$	0.19	±	0.12	0.55	±	0.38	1.26	±	0.80	
	Eos#	$\times 10^9/L$	0.12	±	0.30	0.35	±	0.50	0.80	±	1.20	
	Bas#	$\times 10^9/L$	0.04	±	0.15	0.12	±	0.50	0.22	±	1.00	
	IMG#*	$\times 10^9/L$	0.04	±	0.20	0.11	±	0.40	0.26	±	1.20	
	Neu%	%	64.9	±	12.0	65.4	±	10.0	65.3	±	10.0	
	Lym%	%	23.5	±	9.5	23.0	±	10.0	23.3	±	8.0	
	Mon%	%	6.3	±	3.0	6.2	±	4.0	6.3	±	3.5	
	Eos%	%	3.9	±	6.0	4.0	±	6.0	4.0	±	6.0	
	Bas%	%	1.4	±	5.0	1.4	±	5.0	1.1	±	5.0	
	IMG%*	%	1.3	±	5.0	1.3	±	5.0	1.3	±	5.0	
	RBC	$\times 10^{12}/L$	2.69	±	0.15	4.54	±	0.24	5.34	±	0.30	
	HGB	g/L	84	±	4	142	±	5	173	±	6	
	HCT	%	25.1	±	2.0	43.4	±	2.5	53.6	±	3.0	
	MCV	fL	93.2	±	5.0	95.5	±	5.0	100.4	±	5.0	
	MCH	pg	31.2	±	2.5	31.3	±	2.5	32.4	±	2.5	
	MCHC	g/L	335	±	30	328	±	30	323	±	30	
	RDW-CV	%	15.2	±	5.0	14.9	±	5.0	14.5	±	6.0	
	RDW-SD	fL	52.9	±	10.0	52.6	±	10.0	53.4	±	12.0	
	PLT	$\times 10^9/L$	97	±	20	237	±	40	497	±	60	
	MPV	fL	9.9	±	3.0	9.9	±	3.0	10.0	±	3.0	
	PDW	/	13.8	±	5.0	13.8	±	5.0	13.8	±	5.0	
	PCT	%	0.096	±	0.030	0.235	±	0.070	0.497	±	0.110	
	P-LCR	%	15.4	±	10.0	17.2	±	10.0	17.8	±	10.0	
	P-LCC	$\times 10^9/L$	15	±	8	41	±	24	88	±	50	
	IPF**	%	4.2	±	15.0	4.3	±	15.0	4.2	±	15.0	
	Ret#****	$\times 10^{12}/L$	0.0293	±	0.0230	0.1180	±	0.0990	0.2232	±	0.1200	
	Ret%****	%	1.09	±	0.50	2.60	±	2.00	4.18	±	3.00	
	IRF****	%	9.2	±	15.0	10.4	±	13.0	10.0	±	20.0	
	LFR****	%	90.8	±	15.0	89.6	±	13.0	90.0	±	20.0	
	MFR****	%	7.2	±	13.0	8.4	±	11.0	8.0	±	18.0	
	HFR****	%	2.0	±	2.0	2.0	±	2.0	2.0	±	2.0	
	RHE***	pg	25.6	±	5.0	24.7	±	5.0	25.2	±	5.0	
	NRBC#	$\times 10^9/L$	0.118	±	0.090	0.342	±	0.270	0.796	±	0.600	
	NRBC%	/100WBC	3.97	±	4.00	3.88	±	4.00	3.98	±	4.00	

LOT BC61024

2025-03-10

Instrument	Parameter	Unit	Low Level			Normal Level			High Level		
			LOT	BC61024L-1		LOT	BC61024N-1		LOT	BC61024H-1	
				Values	Range		Values	Range		Values	Range
BC-6600 BC-6800	WBC	$\times 10^9/L$	2.94	±	0.80	8.69	±	1.00	19.69	±	2.00
	Neu#	$\times 10^9/L$	1.91	±	0.50	5.71	±	0.80	12.85	±	2.00
	Lym#	$\times 10^9/L$	0.69	±	0.40	2.01	±	0.80	4.65	±	1.60
	Mon#	$\times 10^9/L$	0.19	±	0.12	0.53	±	0.38	1.22	±	0.80
	Eos#	$\times 10^9/L$	0.12	±	0.30	0.35	±	0.50	0.77	±	1.20
	Bas#	$\times 10^9/L$	0.03	±	0.15	0.09	±	0.50	0.20	±	1.00
	IMG#*	$\times 10^9/L$	0.04	±	0.20	0.10	±	0.40	0.24	±	1.20
	Neu%	%	65.4	±	12.0	65.8	±	10.0	65.3	±	10.0
	Lym%	%	23.3	±	9.5	23.1	±	10.0	23.6	±	8.0
	Mon%	%	6.3	±	3.0	6.1	±	4.0	6.2	±	3.5
	Eos%	%	4.0	±	6.0	4.0	±	6.0	3.9	±	6.0
	Bas%	%	1.0	±	5.0	1.0	±	5.0	1.0	±	5.0
	IMG%*	%	1.3	±	5.0	1.2	±	5.0	1.2	±	5.0
	RBC	$\times 10^{12}/L$	2.74	±	0.15	4.52	±	0.24	5.24	±	0.30
	HGB	g/L	84	±	4	140	±	5	170	±	6
	HCT	%	25.3	±	2.0	42.5	±	2.5	51.8	±	3.0
	MCV	fL	92.2	±	5.0	94.0	±	5.0	98.8	±	5.0
	MCH	pg	30.7	±	2.5	31.0	±	2.5	32.4	±	2.5
	MCHC	g/L	333	±	30	330	±	30	328	±	30
	RDW-CV	%	14.7	±	5.0	14.9	±	5.0	14.7	±	6.0
	RDW-SD	fL	49.1	±	10.0	50.5	±	10.0	52.1	±	12.0
	PLT	$\times 10^9/L$	100	±	20	243	±	40	504	±	60
	MPV	fL	9.4	±	3.0	9.4	±	3.0	9.4	±	3.0
	PDW	/	13.8	±	5.0	13.9	±	5.0	13.8	±	5.0
	PCT	%	0.094	±	0.030	0.228	±	0.070	0.474	±	0.110
	P-LCR	%	11.9	±	10.0	13.5	±	10.0	13.3	±	10.0
	P-LCC	$\times 10^9/L$	12	±	8	33	±	24	67	±	50
	IPF**	%	4.2	±	15.0	4.2	±	15.0	4.2	±	15.0
	Ret#****	$\times 10^{12}/L$	0.0225	±	0.0230	0.1098	±	0.0990	0.2023	±	0.1200
	Ret%****	%	0.82	±	0.50	2.43	±	2.00	3.86	±	3.00
	IRF****	%	6.9	±	15.0	7.1	±	13.0	6.9	±	20.0
	LFR****	%	93.1	±	15.0	92.9	±	13.0	93.1	±	20.0
	MFR****	%	4.5	±	13.0	4.7	±	11.0	4.5	±	18.0
	HFR****	%	2.4	±	2.0	2.4	±	2.0	2.4	±	2.0
	RHE***	pg	28.4	±	5.0	28.6	±	5.0	29.0	±	5.0
	NRBC#	$\times 10^9/L$	0.118	±	0.090	0.340	±	0.270	0.760	±	0.600
	NRBC%	/100WBC	4.00	±	4.00	3.91	±	4.00	3.86	±	4.00

LOT BC61024

2025-03-10

Instrument	Parameter	Unit	Low Level			Normal Level			High Level			
			LOT		BC61024L-1		LOT		BC61024N-1		LOT	
			Values	Range	Values	Range	Values	Range	Values	Range	Values	Range
BC-6800Plus BC-6700Plus BC-6600Plus	WBC	$\times 10^9/L$	3.19	±	0.80	9.06	±	1.00	20.62	±	2.00	
	Neu#	$\times 10^9/L$	2.09	±	0.50	5.89	±	0.80	13.42	±	2.00	
	Lym#	$\times 10^9/L$	0.74	±	0.40	2.12	±	0.80	4.87	±	1.60	
	Mon#	$\times 10^9/L$	0.20	±	0.12	0.56	±	0.38	1.28	±	0.80	
	Eos#	$\times 10^9/L$	0.12	±	0.30	0.36	±	0.50	0.82	±	1.20	
	Bas#	$\times 10^9/L$	0.04	±	0.15	0.13	±	0.50	0.23	±	1.00	
	IMG#*	$\times 10^9/L$	0.04	±	0.20	0.11	±	0.40	0.27	±	1.20	
	Neu%	%	65.1	±	12.0	65.0	±	10.0	65.1	±	10.0	
	Lym%	%	23.3	±	9.5	23.4	±	10.0	23.6	±	8.0	
	Mon%	%	6.3	±	3.0	6.2	±	4.0	6.2	±	3.5	
	Eos%	%	3.9	±	6.0	4.0	±	6.0	4.0	±	6.0	
	Bas%	%	1.4	±	5.0	1.4	±	5.0	1.1	±	5.0	
	IMG%*	%	1.3	±	5.0	1.2	±	5.0	1.3	±	5.0	
	RBC	$\times 10^{12}/L$	2.71	±	0.15	4.50	±	0.24	5.23	±	0.30	
	HGB	g/L	85	±	4	139	±	5	169	±	6	
	HCT	%	25.0	±	2.0	42.4	±	2.5	51.8	±	3.0	
	MCV	fL	92.1	±	5.0	94.2	±	5.0	99.1	±	5.0	
	MCH	pg	31.4	±	2.5	30.9	±	2.5	32.3	±	2.5	
	MCHC	g/L	341	±	30	328	±	30	326	±	30	
	RDW-CV	%	14.7	±	5.0	14.8	±	5.0	14.6	±	6.0	
	RDW-SD	fL	49.1	±	10.0	50.6	±	10.0	52.1	±	12.0	
	PLT	$\times 10^9/L$	95	±	20	239	±	40	496	±	60	
	PLT-H ^{MM}	$\times 10^9/L$	94	±	30	237	±	60	494	±	120	
	PLT-O ^A	$\times 10^9/L$	102	±	30	255	±	60	552	±	120	
	MPV	fL	9.5	±	3.0	9.5	±	3.0	9.4	±	3.0	
	PDW	/	13.9	±	5.0	13.9	±	5.0	13.8	±	5.0	
	PCT	%	0.090	±	0.030	0.227	±	0.070	0.466	±	0.110	
	P-LCR	%	12.6	±	10.0	13.6	±	10.0	13.4	±	10.0	
	P-LCC	$\times 10^9/L$	12	±	8	33	±	24	66	±	50	
	IPF**	%	4.2	±	15.0	4.2	±	15.0	4.3	±	15.0	
	Ret#****	$\times 10^{12}/L$	0.0309	±	0.0230	0.1224	±	0.0990	0.2249	±	0.1200	
	Ret%****	%	1.14	±	0.50	2.72	±	2.00	4.30	±	3.00	
	IRF****	%	8.9	±	15.0	11.5	±	13.0	11.3	±	20.0	
	LFR****	%	91.1	±	15.0	88.5	±	13.0	88.7	±	20.0	
	MFR****	%	7.8	±	13.0	9.5	±	11.0	9.3	±	18.0	
	HFR****	%	1.1	±	2.0	2.0	±	2.0	2.0	±	2.0	
	RHE***	pg	26.4	±	5.0	25.5	±	5.0	26.0	±	5.0	
	NRBC#	$\times 10^9/L$	0.125	±	0.090	0.349	±	0.270	0.779	±	0.600	
	NRBC%	/100WBC	3.92	±	4.00	3.85	±	4.00	3.78	±	4.00	
	MCVr ^{MM}	fL	113.2	±	19.0	114.6	±	13.0	118.0	±	13.0	
	MCHr ^{MM}	pg	31.0	±	6.3	31.4	±	6.3	32.4	±	6.3	
	HDW ^{MM}	g/L	24.9	±	14.0	24.6	±	14.0	24.7	±	14.0	
	DELTA-CH ^{MM}	pg	6.7	±	4.8	6.8	±	4.8	7.0	±	4.8	
	HYPO-CH% ^{MM}	%	2.2	±	12.0	2.2	±	14.0	2.2	±	14.0	
	HYPER-CH% ^{MM}	%	4.8	±	3.7	4.5	±	4.0	4.3	±	4.7	
	MCHm ^{MM}	pg	24.2	±	5.0	24.6	±	5.0	25.4	±	5.0	
	IPF# ^{MM}	$\times 10^9/L$	4.0	±	8.0	10.2	±	25.0	21.6	±	40.0	
	Micro% ^{MM}	%	2.2	±	7.0	2.0	±	3.0	2.0	±	5.0	
	Macro% ^{MM}	%	5.0	±	8.0	5.1	±	10.0	5.9	±	10.0	
	Neu-X ^{MM}	/	447.3	±	120.0	453.5	±	120.0	459.1	±	120.0	
	Neu-Y ^{MM}	/	447.5	±	170.0	447.4	±	170.0	448.7	±	170.0	
	Mon-X ^{MM}	/	268.9	±	120.0	263.2	±	120.0	263.7	±	120.0	
	PLT-I ^{MM}	$\times 10^9/L$	96	±	30	244	±	60	506	±	120	

LOT BC61024

2025-03-10

Instrument	Parameter	Unit	Low Level			Normal Level			High Level			
			LOT		BC61024L-1		LOT		BC61024N-1		LOT	
			Values	Range	Values	Range	Values	Range	Values	Range	Values	Range
BC-6000 BC-6100 BC-6200 BC-6000Plus BC-6100Plus	WBC	$\times 10^9/L$	2.96	±	0.80	8.88	±	1.00	20.21	±	2.00	
	Neu#	$\times 10^9/L$	1.91	±	0.50	5.79	±	0.80	13.16	±	2.00	
	Lym#	$\times 10^9/L$	0.70	±	0.40	2.06	±	0.80	4.75	±	1.60	
	Mon#	$\times 10^9/L$	0.19	±	0.12	0.55	±	0.38	1.27	±	0.80	
	Eos#	$\times 10^9/L$	0.12	±	0.30	0.36	±	0.50	0.81	±	1.20	
	Bas#	$\times 10^9/L$	0.04	±	0.15	0.12	±	0.50	0.22	±	1.00	
	IMG#*	$\times 10^9/L$	0.04	±	0.20	0.12	±	0.40	0.26	±	1.20	
	Neu%	%	64.7	±	12.0	65.2	±	10.0	65.1	±	10.0	
	Lym%	%	23.6	±	9.5	23.2	±	10.0	23.5	±	8.0	
	Mon%	%	6.4	±	3.0	6.2	±	4.0	6.3	±	3.5	
	Eos%	%	3.9	±	6.0	4.0	±	6.0	4.0	±	6.0	
	Bas%	%	1.4	±	5.0	1.4	±	5.0	1.1	±	5.0	
	IMG%*	%	1.3	±	5.0	1.3	±	5.0	1.3	±	5.0	
	RBC	$\times 10^{12}/L$	2.73	±	0.15	4.51	±	0.24	5.27	±	0.30	
	HGB	g/L	85	±	4	141	±	5	171	±	6	
	HCT	%	25.7	±	2.0	42.9	±	2.5	52.3	±	3.0	
	MCV	fL	94.2	±	5.0	95.1	±	5.0	99.3	±	5.0	
	MCH	pg	31.1	±	2.5	31.3	±	2.5	32.4	±	2.5	
	MCHC	g/L	331	±	30	329	±	30	327	±	30	
	RDW-CV	%	15.4	±	5.0	15.3	±	5.0	14.7	±	6.0	
	RDW-SD	fL	52.3	±	10.0	52.7	±	10.0	52.9	±	12.0	
	PLT	$\times 10^9/L$	103	±	20	250	±	40	541	±	60	
	PLT-O^	$\times 10^9/L$	99	±	30	267	±	60	601	±	120	
	MPV	fL	10.5	±	3.0	10.3	±	3.0	10.2	±	3.0	
	PDW	/	14.0	±	5.0	14.1	±	5.0	14.1	±	5.0	
	PCT	%	0.108	±	0.030	0.258	±	0.070	0.552	±	0.110	
	P-LCR	%	20.0	±	10.0	20.1	±	10.0	19.5	±	10.0	
	P-LCC	$\times 10^9/L$	21	±	8	50	±	24	105	±	50	
	IPF**	%	4.2	±	15.0	4.2	±	15.0	4.2	±	15.0	
	Ret#****	$\times 10^{12}/L$	0.0303	±	0.0230	0.1159	±	0.0990	0.2161	±	0.1200	
	Ret%****	%	1.11	±	0.50	2.57	±	2.00	4.10	±	3.00	
	IRF****	%	10.8	±	15.0	12.2	±	13.0	12.1	±	20.0	
	LFR****	%	89.2	±	15.0	87.8	±	13.0	87.9	±	20.0	
	MFR****	%	8.8	±	13.0	10.2	±	11.0	10.1	±	18.0	
	HFR****	%	2.0	±	2.0	2.0	±	2.0	2.0	±	2.0	
	RHE***	pg	26.1	±	5.0	24.9	±	5.0	25.4	±	5.0	
	NRBC#	$\times 10^9/L$	0.119	±	0.090	0.342	±	0.270	0.776	±	0.600	
	NRBC%	/100WBC	4.01	±	4.00	3.85	±	4.00	3.84	±	4.00	
	MCVr ^{MM}	fL	111.3	±	19.0	112.0	±	13.0	115.4	±	13.0	
	MCHr ^{MM}	pg	30.4	±	6.3	30.7	±	6.3	31.6	±	6.3	
	HDW ^{MM}	g/L	24.8	±	14.0	24.4	±	14.0	24.4	±	14.0	
	DELTA-CH ^{MM}	pg	6.6	±	4.8	6.7	±	4.8	6.9	±	4.8	
	HYPO-CH% ^{MM}	%	2.2	±	12.0	2.2	±	14.0	2.3	±	14.0	
	HYPER-CH% ^{MM}	%	4.7	±	3.7	4.5	±	4.0	4.3	±	4.7	
	MCHm ^{MM}	pg	23.8	±	5.0	24.0	±	5.0	24.8	±	5.0	
	IPF# ^{MM}	$\times 10^9/L$	4.3	±	8.0	10.5	±	25.0	22.7	±	40.0	
	Micro% ^{MM}	%	2.3	±	7.0	2.2	±	3.0	2.0	±	5.0	
	Macro% ^{MM}	%	5.9	±	8.0	6.0	±	10.0	6.9	±	10.0	
	Neu-X ^{MM}	/	463.8	±	120.0	461.8	±	120.0	457.9	±	120.0	
	Neu-Y ^{MM}	/	449.4	±	170.0	452.0	±	170.0	451.6	±	170.0	
	Mon-X ^{MM}	/	292.0	±	120.0	291.4	±	120.0	289.7	±	120.0	
	PLT-I ^{MM}	$\times 10^9/L$	103	±	30	250	±	60	541	±	120	

LOT BC61024

2025-03-10

Instrument	Parameter	Unit	Low Level			Normal Level			High Level		
			LOT	BC61024L-1		LOT	BC61024N-1		LOT	BC61024H-1	
				Values	Range		Values	Range		Values	Range
BC-700[B] BC-760[B] BC-700[B] CS BC-760[B] CS	WBC	$\times 10^9/L$	2.99	±	0.80	9.02	±	1.00	20.39	±	2.00
	Neu#	$\times 10^9/L$	2.01	±	0.50	6.07	±	0.80	13.70	±	2.00
	Lym#	$\times 10^9/L$	0.68	±	0.40	2.05	±	0.80	4.67	±	1.60
	Mon#	$\times 10^9/L$	0.19	±	0.12	0.56	±	0.38	1.26	±	0.80
	Eos#	$\times 10^9/L$	0.07	±	0.30	0.21	±	0.50	0.47	±	1.20
	Bas#	$\times 10^9/L$	0.04	±	0.15	0.13	±	0.50	0.29	±	1.00
	IMG#*	$\times 10^9/L$	0.04	±	0.20	0.12	±	0.40	0.27	±	1.20
	Neu%	%	67.0	±	12.0	67.4	±	10.0	67.2	±	10.0
	Lym%	%	22.9	±	9.5	22.7	±	10.0	22.9	±	8.0
	Mon%	%	6.4	±	3.0	6.2	±	4.0	6.2	±	3.5
	Eos%	%	2.4	±	6.0	2.3	±	6.0	2.3	±	6.0
	Bas%	%	1.3	±	5.0	1.4	±	5.0	1.4	±	5.0
	IMG%*	%	1.3	±	5.0	1.3	±	5.0	1.3	±	5.0
	RBC	$\times 10^{12}/L$	2.70	±	0.15	4.57	±	0.24	5.34	±	0.30
	HGB	g/L	83	±	4	140	±	5	170	±	6
	HCT	%	25.1	±	2.0	43.6	±	2.5	53.5	±	3.0
	MCV	fL	93.0	±	5.0	95.5	±	5.0	100.2	±	5.0
	MCH	pg	30.7	±	2.5	30.6	±	2.5	31.8	±	2.5
	MCHC	g/L	331	±	30	321	±	30	318	±	30
	RDW-CV	%	15.4	±	5.0	14.8	±	5.0	14.4	±	6.0
	RDW-SD	fL	53.9	±	10.0	53.0	±	10.0	53.4	±	12.0
	PLT	$\times 10^9/L$	102	±	20	241	±	40	512	±	60
	PLT-H ^{MM}	$\times 10^9/L$	103	±	30	244	±	60	519	±	120
	MPV	fL	9.8	±	3.0	9.9	±	3.0	9.8	±	3.0
	PDW	/	13.7	±	5.0	13.7	±	5.0	13.8	±	5.0
	PCT	%	0.100	±	0.030	0.239	±	0.070	0.502	±	0.110
	P-LCR	%	14.3	±	10.0	15.7	±	10.0	16.4	±	10.0
	P-LCC	$\times 10^9/L$	15	±	8	38	±	24	84	±	50
	IPF**	%	4.2	±	15.0	4.2	±	15.0	4.2	±	15.0
	NRBC#	$\times 10^9/L$	0.056	±	0.090	0.169	±	0.270	0.375	±	0.600
	NRBC%	/100WBC	1.86	±	4.00	1.87	±	4.00	1.84	±	4.00

*Applies to BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[N] CRP, BC-7500[NR] CRP, BC-7300[N] CRP, BC-7300[NR] CRP, BC-7500[N] CS, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS and Software version 1.09 or higher of BC-6600/BC-6800.

**Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS and software version 1.09 or higher of BC-6600/BC-6800.

***Applies to BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS and Software version 1.10 or higher of BC-6800, BC-6600.

****Applies to BC-6600, BC-6800, BC-6200, BC-6000Plus, BC-6100Plus, BC-6800Plus, BC-6700Plus, BC-6600Plus, BC-7500[NR] CRP, BC-7300[NR] CRP, BC-7500[NR] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

^Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS.

^^Applies to BC-700[B], BC-760[B], BC-700[B] CS, BC-760[B] CS, BC-700[R], BC-720[R], BC-760[R], BC-780[R], BC-700[R] CS, BC-760[R] CS, BC-6600Plus, BC-6700Plus.

^^^Applies to BC-6000, BC-6100, BC-6000Plus, BC-6100Plus, BC-6200, BC-6600Plus, BC-6700Plus, BC-6800Plus.