

Hemoglobin A1c Control

mindray

Value sheet of Mindray Automated Glycohemoglobin Analyzer

LOT:202305102		:2024-05-09					
Parameter	Method	Analyzer	Level	Unit		Assay Value	Range
HbA1c	HPLC	H50/H50P	CRL-1	IFCC	mmol/mol	32	26~38
				NGSP	%	5.1	4.6~5.6
			CRL-2	IFCC	mmol/mol	95	87~102
				NGSP	%	10.8	10.2~11.4